



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1,
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Keli Electric Manufacturing (Ningbo) Co., Ltd.
Address: 199 Changxing Road, Jiangbei District,
Ningbo City,
P.R. China

Manufacturer of the certified type

Name: Keli Electric Manufacturing (Ningbo) Co., Ltd.
Address: 199 Changxing Road, Jiangbei District,
Ningbo City,
P.R. China

Identification of certified type

Type : ZSF-A-..... and ZSFY-A-.....
Fraction : $P_i = 0.7$
Temperature range $-10\text{ }^{\circ}\text{C} / +40\text{ }^{\circ}\text{C}$

Maximum capacity (E_{\max})	10 000 kg up to and including 50 000 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Ratio of minimum LC Verification interval $Y = E_{\max} / V_{\min}$	10 000



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R60/2000-NL1-05.17 revision 1
Project number 507729
Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC6827 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60
Edition 2000 (E)
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R60/2000-NL1-05.17A that includes 40 pages.

N° R60/2000-NL1-05.17B that includes 37 pages.

This revision replaces the earlier version.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


1 November 2005

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.